



गंगादेव गाउँपालिका  
गाउँ कार्यपालिकाको कार्यालय  
सानडाँडा, रोल्पा



पत्र संख्या: २०८३/०८३

सूचना नं: २४



लुम्बिनी प्रदेश, नेपाल

स्वास्थ्य सामग्रीहरू तथा स्टेचरको दररेट उपलब्ध गराईदिने सम्बन्धी सूचना

मिति: २०८३/०२/०९ (ने.सं. ११४६ बछलागा त्रयोदशी शुक्रबार)

यस गंगादेव गाउँपालिका अन्तर्गतका स्वास्थ्य संस्थाहरूमा आर्थिक वर्ष २०८२/०८३ का लागि स्वास्थ्य सर्जिकल सामग्रीहरू र स्टेचर खरिद गरी वितरण गर्नुपर्ने भएको हुँदा यसै सूचना साथ संलग्न सूचि अनुसारको ISO/USFDA Approved Product Certified सामग्रीहरूको प्रचलित दररेट सूचना प्रकाशन भएको मितिले ७ (सात) दिनभित्र यस गाउँपालिकाको कार्यालयमा आवश्यक कागजात सहित सप्लायर्स स्वयंम् आफै उपस्थित भई वा कार्यालयको ई-मेल (gangadevmun@gmail.com) मार्फत उपलब्ध गराई दिनुहुन सम्बन्धित सबैको लागि यो सूचना प्रकाशित गरिएको छ।


आवश्यक कागजातहरू:

१. संस्था/सप्लायर्स/फर्म दर्ता तथा नविकरणको प्रमाण-पत्र।
२. स्थायी लेखा नं. (PAN/VAT) को प्रमाण-पत्र।
३. कर चुक्ता प्रमाण-पत्र।
४. सम्बन्धित व्यावसायको इजाजत पत्र।

याम बहादुर बुढामगर  
प्रमुख प्रशासकीय अधिकृत

तपसीलका स्वास्थ्य संस्थाहरूमा आवश्यक स्वास्थ्य सामग्रीहरूको विवरणः

SN	Name of medical instrument	Specicication	Unit	Quantity	Unit Price	If vatable,13% add unit rate	Total unit rate(Unit rate+ 13% VAT)	Remark
1	Peri light	1)light weight and easy to carry/ move (2) lamp position adjustable up to 40 degree backward (3) lamp power 50 watt or more with extra focus (4) On/Off switch facility (5) Focus adjustment for specific examination area (6) Long lifespan lamp (7) Flexible mobile stad with caster wheels (8) Height adjustable stand ((9)	Piece	1				
2	Nebulizer machine	(1)light weight, compact, portable dsgn and easy to carry (2) Type- Compressor nebulizer (3) Partical size- MMAD approximately 0.5 to 5 microns (4) On/Off switch facility (5) Nebulization rate- minimum 0.2 ml/ minutes (6) Medication cup capacity- 6 to 10 ml (7) Residual medication volume- less than 1 ml preferred (8) Air flow rate- Minimum 5 to 8 litre per minutes (9) Operating pressure - approximately 8 to 16 psi (10) Noise level- less than 60 dB preferred (11) continuous operation capability .	Set	1				
3	Dressing Trolley	(1)Heavy carriage mounted on 4 swivel anti-static castors, of which two with brakes (2)Fit on both sides with push bar-handle. (3)Top and bottom shelves with guard rails. (4)Protective bumpers at all four corners. (5) Frame and tray: Austenitic stainless steel 18/10. (6)Dimensions: <ul style="list-style-type: none"> <li>• Overall: Approximately (+ 10%) 900 x 550 x 1000mm (l x w x h).</li> <li>• Frame, diameter: 30mm.</li> <li>• Thickness of shelves: Approximately 1.5mm</li> <li>• Swivel castors approx. diameter: 100mm.</li> <li>• Carrying capacity: Approx. 100kg.</li> </ul>	Than	1				
4	Medicine Trolley	(1) It shall be constructed fully with 304 grade stainless steel sheet and tube or better. (2) Overall size: approximately 900 H x 460 W x 760 L mm. (3) Flat top of SS and at least 6 inch deep removable bucket at bottom. (4) Multiple drawers (minimum 6) made of high quality materials with telescopic channels, below the platform. (5) Minimum 6 small boxes to keep drugs at eye level. (6)Equipped with waste bin, needle disposable container, file cassette, atraumatic corner buffers and guard rails. (7) Shall have provision for hanging one IV fluid bottle. (8) Equipped with lock key system. (9) Shall be mobile on 4 x 100mm diameter (approx.) robust 360 deg. anti-rust, anti-static, noiseless, swivel castors with non-marking grey tyres and with at least 2 diagonal castors shall have brakes.	Than	1				
5	Examination Bed	Examination Bed for patient	Than	1				

  
 सिकलदेव  
 स्वास्थ्य




6	Iv Stand	(1) Adjustable IV/saline stand with five legs, with double hooks and five swivels castors. (2) The IV stand shall be made of tubular anti-corrosive and antirust treated epoxy powder coated mild steel, with a 5 pronged base fitted on mobile on swivelling castors of approx. diameter 50mm. The castors must be non-rusting and non-marking. (3) The stand should come with stainless steel double IV hook, height adjustable from approximately 1620mm to 2340mm, with a screw knob for height adjustment.							
7	Dight Baby weaing Scale	Dight Baby weaing Scale	Piece	1					
8	Adult Hight Board	Adult Hight Board	Piece	1					
9	Episiotomy set	(1) Episiotomy Scissors – Braun-Stadler (one each of 14.5cm & 22 cm). -2 (2) Needle Holder (Mayo-Hegar) 20 cm. -1 (3)Dissecting Forceps – Toothed, Oehler, 14cm.-1 (4) Stitch Cutting Scissors - (Abdominal, Kelly) 18cm. -1 (5) Triangular Cutting Needle. -1 (6) Round Body Needle. -1 (7) Sponge Holding Forceps (Forester; Straight; serrated) 25cm. -1 (8)Small Bowl SS – 750ml. -1 (9) Galipot SS 6 ounces 100X50 mm -1 (10)All instruments must be supplied free of residual scale, acid, grease and grinding and polishing materials and workmanship must be first class throughout. Instruments must be free of defects, which detract from their appearance or impair serviceability, proper functioning and intended use.	Piece	1					
10	Cervical tear repair set	Artery Forceps Straight( Haemostatic , Rankin -Crile or Rochester-Pean ) 20 cm -1 , Artery Forceps Curve( Haemostatic , Rankin -Crile or Rochester-Pean ) 20 cm- 1,Sponge Holding Forceps ( Forester, Straight, Serrated ) 24 cm -4, Bowl S. S. (Small) -1, Needle Holder ( Mayo Hegar ) 20 cm -1, Scissors ( Abdominal , Kelly ) 18 cm -1, Suture Cutting Scissors ( Abdominal , Kelly ) long -1 ,Dissecting Forceps toothed , Potts-Smith, 15 cm -1 , Dissecting Forceps Non-toothed , Potts-Smith, 15 cm -1 , Cuscos's vaginal Spaculum -1 , Vulsellum forceps -1 , Kidney tray small size -1, Galli pot -1 , Instrument tray with cover (300 x 200 x 65)mm. -1, Sims Vaginal Spaculum in different size -3	Set	1					
11	BP instruments	(1) Aneroid splygmomanometer (2) Cuffs for adult size (regular) (3) Inflation bulb (4) Carrying pouch (5) Packed in easy carrying high quality pouch made of waterproof cloth to accommodate cuff, and inflation bulb. (6)Gauge to be calibrated in 2 mm Hg units. (7) Must provide blood pressure cuffs for child size and for adult size (regular).	Set	1					
12	Stethoscope	(1) Dual, cup/bell and diaphragm head (2)Head and ear tube assembly to be made of non-ferrous metal, (3)Tubes to be synthetic material and ear tubes to have shaped plastic cushion ends	Piece	1					
13	Fetus doppler Heart Detector With Rechargeble battery	Fetus doppler Heart Detector With Rechargeble battery	Piece	1					
14	Fetus doppler Heart Detector	Fetus doppler Heart Detector	Set	1					
15	Digital pan Weighing Machine	Digital pan Weighing Machine	Set	1					
16	Bedside locker	Bedside locker	Than	1					
17	2 in 1 weight machine	2 in 1 weight machine	Piece	1					

सकलदेव यादव  
स्वास्थ्य संयोजक




18	Resuscitation Set	(1) Mask-The mask or facemask is connected to the bag and covers the mouth and nose of the newborn who is being resuscitated (2) Size 0 for preterm and low-weight baby, round type, diameter 35-50 mm (3) • Size 1 for term baby, round type, diameter 50-65 mm translucent (4) Self-inflating neonatal resuscitator bag, Hand operated, portable, reusable Translucent and made from silicone rubber, capacity approx. 220 – 320ml to provide positive pressure to pre term babies, low birth weight and term babies < 5kg (5) Valve- to prevent excessive pressure, which can cause lung damage, with a pressure limiting valve so that the airway pressure does not exceed 4.5 kPa (45 cmH2 O) and can generate an airway pressure of at least 3 kPa (30 cmH2 O)	Set	1				
19	Room heater	Room heater	Piece	1				
20	MVA set	MVA set existing of 1 Plastic Aspirator and 8 Polyethylene aspirator cannulae of size 4, 5, 6, 7, 8, 9, 10, and 12 mm (colour coded)	Set	1				
21	Suture set	(1) Artery Forceps (Haemostatic, Rankin-Crile or Rochester-Pean) 16cm. -2 (2) Stitch Cutting Scissors - (Abdominal, Kelly) 18cm. -1 (3) Sponge Holding Forceps (Forester; Straight; serrated) 20cm. -1 (4) Bowl S. S. (Small) -1 (5) Needle Holder (Mayo-Hegar) 20 cm. -1 (6) Dissecting Forceps – Toothed, Oehler, 14cm. -1 (7) Dissecting Forceps –Non- Toothed, Oehler, 14cm. -1 (8) Kidney Tray 400mm -1 (9) All instruments must be supplied free of residual scale, acid, grease and grinding and polishing materials and workmanship must be first class throughout. Instruments must be free of defects, which detract from their appearance or impair serviceability, proper functioning and intended use.	Set	1				
22	Dressing Set	(1) Artery Forceps (Haemostatic, Rankin-Crile or Rochester-Pean) 16cm. -2 (2) Stitch Cutting Scissors - (Abdominal, Kelly) 18cm. -1 (3) Sponge Holding Forceps (Forester; Straight; serrated) 20cm. -1 (4) Bowl S. S. (Small) -1 (5) Needle Holder (Mayo-Hegar) 20 cm. -1 (6) Dissecting Forceps – Toothed, Oehler, 14cm. -1 (7) Dissecting Forceps –Non- Toothed, Oehler, 14cm. -1 (8) Kidney Tray 400mm -1 (9) All instruments must be supplied free of residual scale, acid, grease and grinding and polishing materials and workmanship must be first class throughout. Instruments must be free of defects, which detract from their appearance or impair serviceability, proper functioning and intended use.	Set	1				
23	proctoscope	(1) Made of high quality surgical grade stainless steel. (2) Corrosion resistant rust proof and non-magnetic preferred. (3) Complete proctoscope set consisting of; outer tube/sheath, Obturator, Handle and light carrier adapter (4) Smooth round distal end for atraumatic insertion (5) Dimensions; tube diameter approximately 20-25 mm, working length approximately 80-150 mm (6) Fully autoclavable at standard hospital sterilization conditions.	Set	1				

  
संस्कारदाव यादव  
स्वास्थ्य संयोजक

  
गंगादेव गाउँपालिका  
गाउँ कार्यपालिकाको कार्यालय  
स्वास्थ्य शाखा  
साविक-०५, रोल्पा  
लुम्बिनी प्रदेश, नेपाल  
२०७३

24	ILR freeze Single compartment (195-220 litres)	voltage regulator/stabilizer. (2)Vaccine storage capacity: approx. 90-130 litres with baskets in place. (3)Construction: • Corrosion resistance: Internal and external cabinet, lid and frame protected against corrosion to DIN8985. Bidder to specify the materials used for construction of internal & external cabinet. • Chest type with CFC – free insulation. • Solid door with lock and handle. (4)Type: Compression Cycled, CFC-Free Refrigerant R-134a (both for refrigeration and insulation) Cooling coil of copper.(5)Holdover time: The refrigerator shall maintain +2 oC to +8 oC for a period as per WHO PQS requirements and preferably higher hours during power cut at the ambient temperature of +43 oC. Bidder shall provide details of holdover time of their product. (6)Compressor: • Sealed heavy duty type. • Starts and operates on electricity supply from 165 to 255 volts. (7)Control Panel: The thermostat, thermometer and other visual displays like On/Off switch, power indicator must be positioned: • on the front of the unit, as close to eye level as possible or • on top of the unit at a height not exceeding 1.3 metres or • If a low level position is essential, the display must be aligned so that it can easily be read without the user having to squat or kneel down. (8)Shall have adjustments for uneven bases. The adjustments must be easy to use like rotating a screw at the legs in the base.(9)On/Off switch and power indicator	Than	1				
25	Needle Cutter Electric	(1)Size – not more than 18cm length, 13cm width & 13 cm height. (2)Weight not more than 4kg. (3) Manual cutter – hardened blade of stainless material (4) Needle burning capacity – to destroy Inj. Needles of length 12.5mm to 80 mm. (5) A needle of 1.6mm diameter & 80 mm length should be destroyed in 1 minute.	Than	1				
26	Bedside tools or Bench for visitors	(1)Bedside tools or Bench for visitors (2) steel frame, minium 30 cm diameter, 45-50 cm height, 4 legs stable structure , light weight and easy to carry (3) 100 kg weight bearing capacity	Piece	1				
27	Cheatle forcep with jar	(1) Material: Stainless steel, autoclaveable. (2) Conical shaped with broad base for keeping cheatle forceps of size 270 mm.	Set	1				
28	knee hammer	(1) Type : Taylor/Buck/Tromner or equivalent. (2) Head made of soft rubber. (3) Handle made of stainless steel/ chrome plated metal or equivalent. (4) Light Weight and balance design. (5) Reusable and easy to disinfect. (6) Standard medical grade length	Piece	1				

  
संकलदेव यादव  
स्वास्थ्य संयोजक




29	infantometer	<p>(1)Measuring range in Cm : Minimum 10 cm – 100 cm  (2)Measuring range in inch: 4 – 39"  (3) Graduation {Measuring Rod} : 1 mm / 1/ 16 inch  (4) Measuring Scale Shall be Permanently marked.  (5) Head board shall be fixed and perpendicular.  (6) Foot piece shall be movable and locked.  (7) Material shall be non-corrosive and easy to clean.  (8)Surface shall be smooth with no Sharp edges.  (9)Struture shall be stable during measurement  (10) Measurement unit ; centimeter and millimeter.  (11)Measure (WxHxD): 1110x115 x 333 mm, 43,7 x 4,5 x 13,1 inch  (12)Net weight : 1,6 kg, 3,5 lbs  (13) Lightweight and portable design  (14) Sturdy and robust infantometer.</p>						
30	adult weight machine	adult weight machine	Than	1				
	Instrumental tray (480*330*20 mm) size		Piece	1				
31		<p>(1)Seamless tray, rectangular with rounded corners, stainless steel, smooth surface.(2)Stain-proof and long lasting. (3) Austenitic stainless steel. (4)Dimensions:  • Length: approximately 300mm.  • Width: approximately 200mm.  • Height: approximately 30mm.  • Thickness: 0.8mm.</p>						
	Instrumental tray (300*200*30 mm) size		Piece	1				
32		<p>(1)Seamless tray, rectangular with rounded corners, stainless steel, smooth surface.(2)Stain-proof and long lasting. (3) Austenitic stainless steel. (4)Dimensions:  • Length: approximately 300mm.  • Width: approximately 200mm.  • Height: approximately 30mm.  • Thickness: 0.8mm.</p>						
	Instrumental tray (310*195*60 mm) size		Piece	1				
33		<p>(1)Seamless tray, rectangular with rounded corners, stainless steel, smooth surface.(2)Stain-proof and long lasting. (3) Austenitic stainless steel. (4)Dimensions:  • Length: approximately 300mm.  • Width: approximately 200mm.  • Height: approximately 30mm.  • Thickness: 0.8mm.</p>						
			Piece	1				

*[Signature]*  
**संकलदेव यादव**  
**स्वास्थ्य संयोजक**


**गंगादेव गाउँपालिका**  
**गाउँ कार्यपालिकाको कार्यालय**  
**स्वास्थ्य शाखा**  
**सान्नीस रोड**  
**लुम्बिनी प्रदेश नेपाल**  
**२०७३**

34	Instrumental tray (225*125*50 mm) size	(1)Seamless tray, rectangular with rounded corners, stainless steel, smooth surface.(2)Stain-proof and long lasting. (3) Austenitic stainless steel. (4)Dimensions: • Length: approximately 300mm. • Width: approximately 200mm. • Height: approximately 30mm. • Thickness: 0.8mm.	Piece	1					
35	Implant Insertion Set	(1) Gallis pot; Galli pot--4 onz, 1.5" height, SS -2 (2) Dissecting forceps, 6", straight, (Non-toothed) -1 (3) Sponge holder, 9.5", straight -1 (4) Implant trochars with canula -1 (5) Syringe, disposable 5 ml with 2" needle -1 (6) Artery Forceps curved 14 cm -1 (7) Eye Piece green cotton cloth tawel -2 (8) Cotton Cloth Double rapper with drawstring -1	Set	1					
36	Implant Removal Set	(1) Gallis pot; Galli pot--4 onz, 1.5" height, SS -2 (2) Dissecting forceps, 6", straight, (Non-toothed) -1 (3) Sponge holder, 9.5", straight -1 (4) Surgical handle no. 3 -1 (5) Mosquito forceps, curved 5" -1 (6) Ringed Forceps for NSV (3.5 mm) -1 (7) Syringe, disposable 5 ml with 2" needle -1 (8) Artery Forceps curved 14 cm -1 (9) Eye Piece green cotton cloth tawel -2 (10) Cotton Cloth Double rapper with drawstring -1	Set	1					
37	IUCD Insertion Set	(1) Gallis pot; Galli pot--4 onz, 1.5" height, SS -1 (2) Kidney Tray, 10" -1 (3) Forceps-artery, Pean, straight, 8 1/2" -1 (4) Volselum, 9.5", curved -1 (5) Sponge holder, 9.5", straight -1 (6) Mayo dissecting scissors s/s curved, 7" -1 (7) Uterine sound, 12.5", metallic -1 (8) Bi-valve Vaginal speculum medium -1 (9) Eye Piece green cotton cloth tawel -2 (10)Cotton Cloth Double rapper with drawstring -1	Set	1					
38	IUCD Removal Set	(1) Gallis pot; Galli pot--4 onz, 1.5" height, SS -1 (2) Kidney Tray, 10" -1 (3) Forceps-artery, Pean, straight, 8 1/2" -1 (4) Volselum, 9.5", curved -1 (5) Sponge holder, 9.5", straight -1 (6) Bi-valve Vaginal speculum medium -1 (7) Eye Piece green cotton cloth tawel -2 (8)Cotton Cloth Double rapper with drawstring -1	Set	1					
39	Penguin suction	(1) medical grade non-toxic silicone made (2) approximately capacity 60-90 ml (3) Soft squeezable bulb with tapered nozzle (4) Leak proof design	Piece	1					
40	Autoclave Vertical 20-25 litre electrical	Autoclave Vertical 20-25 litre electrical	Piece	1					
41	Autoclave Vertical 20-25 litre Non electrical	Autoclave Vertical 20-25 litre Non electrical	Piece	1					
42	Delivery set	(1) Artery Forceps (Haemostatic, Rankin-Crile or Rochester-Pean) 16cm.-2 (2) Cord Cutting Scissors (Umbilical; Blunt end) 16 cm.-1 (3) Sponge Holding Forceps (Forester; Straight; serrated) 20cm.-1 (4) Bowl S. S. (Small) 600 - 750 ml.-1 (5) Gally Pot -1 (6) Baby Wrapper Cotton and Flatin cloths 1*1 ( 4 ) ( Cotton -2, Flatin-2)-4 (7) Delivery Set Single and Double wrapper (8) Rubber Sheet (1*1)-1 (9) Penguin Suction -1 (10) All instruments must be supplied free of residual scale, acid, grease and grinding and polishing materials and workmanship must be first class throughout. Instruments must be free of defects, which detract from their appearance or impair serviceability, proper functioning and intended use.	Set	1					

  
 सुकन्देव यादव  
 स्वास्थ्य संयोजक




43	Neonatal Radiant Baby Warmer with Baby Bassinet	<p>(1) It must have facility to display both skin and air (ambient) temperature separately. (2) In servo mode, the heater output must be controlled to maintain the baby at the required set temperature. (3) It must have inbuilt rechargeable battery to run equipment in case of power failure for at least 30 min. (4) It must rotate and swivel in different direction, so as to allow taking X-rays. (5) It must have manual setting for high and low alarm setting. (6) It shall have audio-visual alarm facilities for:</p> <ul style="list-style-type: none"> <li>• Overheating beyond set temperature range.</li> <li>• Patient temperature less than or greater than the required temperature i.e. above or below the set range. <ul style="list-style-type: none"> <li>• Power failure.</li> <li>• Heater failure.</li> <li>• Probe failure.</li> </ul> </li> <li>• Time out alarm in manual mode.</li> </ul> <p>(7) The light must be dazzle free. (8) In manual mode, the heater output must be directly controlled by a setting on the front panel. (9) The desired temperature range from 25 to 40°C. (10) The resolution must be 0.1°C. (11) Halogen based observation light must be provided for observing the baby (12) It must be mounted on a pole with sturdy base with lockable castors.</p>	Than	1				
44	Standard Delivery bed with water proof mattress and pillow	Standard Delivery bed with water proof mattress and pillow	Than	1				
45	Hospital Bed Plain with matters pillow	Hospital Bed Plain with water proof matters and pillow	Than	1				
46	Semi floding Bed with mattres pillow	Semi floding Bed with mattres pillow	Than	1				
47	Double Foot step	Double Foot step	Than	1				
48	Patient Bed Side Screen	Patient Bed Side Screen 4fold with cotton green cloth	Than	1				
49	Bed Side locker	Bed Side locker	Than	1				
50	Electric Suction Machine	Electric Suction Machine	Set	1				
51	Double Folding Stretcher	Double Folding Stretcher	Than	1				
52	Bed General	Bed General	Than	1				
53	Endo-Tracheal Tube Set	Endo-Tracheal Tube Set	Set	1				
54	Instrument Trolley	Instrument Trolley	Than	1				

  
संकेतितेव यादव  
स्वास्थ्य संयोजक


  
गुगादेव गाउँपालिका  
गाउँ कार्यपालिकाको कार्यालय  
स्वास्थ्य शाखा  
सानेढोडा रोड  
लुम्बिनी प्रदेश  
नेपाल  
२०७३

55	Delivery Bed Hydraulic	<p>coated washable paint at least 12 gauge. (4) Table top must have 3 adjustable parts: (5) Ratchet operated rising backrest, retractable foot end and a fixed centre part. (6) Two support handles shall be positioned on either side of the rising backrest. (7) Leg support stands with waterproof and washable protection.(8)Vertical movement control/adjustment via hydraulically operated monobloc actuator with dual-sided pedal control. (9)All movement and control via gas-spring operated by head/side levellers. (10)Quick release mechanism for back section adjustment during emergency situation for quick passage to CPR (Cardio Pulmonary Resuscitation) position. (11)Shall have provision of IV pole attachment. (12)The bed shall be mounted on four anti-static castor wheels of dia. approx. 100mm, two with brakes (13) Three separate mattresses of at least 100 mm thickness for each section fixed in position to the top of the bed by Velcro strips on the underside of each mattress section, pad covering material to be anti-bacteria, anti-static, waterproof, fire retardant and washable. (14)The Delivery bed shall have maximum weight at least 200Kg. (15)Birth permitted in at least the following positions:</p> <ul style="list-style-type: none"> <li>• Classic Lithotomic.</li> <li>• Sitting and/or semi-sitting.</li> <li>• Knee-elbow.</li> <li>• Vertical.</li> <li>• Squatting.</li> <li>• Caesarean, surgical and foetus extraction.</li> </ul> <p>(16)Head and Lower Ends:</p> <ul style="list-style-type: none"> <li>• Easily removable (for resuscitation).</li> <li>• Made of Stainless Steel with plastic-covered bends and grips.</li> <li>• Removable lower end for easy access to the delivery.</li> </ul>	Than	1				
56	Speculum sims Vaginal Large	<p>(1)Stainless steel, reusable Sim's vaginal speculum, double ended.(2)Material: High grade fully stainless steel, corrosion resistance. (3)Maximum opening of blade 45 degrees (each).(4)Sim's vaginal speculum of three different sizes. (5)Approx. Dimensions:</p> <ul style="list-style-type: none"> <li>• Large: 72x34/80x38mm.</li> </ul> <p>(6)Handle thickness approx. 2.5mm. (7) Workmanship: All surfaces shall be free from burrs, pits, cracks. Edges shall be smoothly rounded off &amp; shall not be sharp.(8)It shall be autoclaveable/sterilizeable.</p>	Piece	1				
57	Speculum sims Vaginal Medium	<p>(1)Stainless steel, reusable Sim's vaginal speculum, double ended.(2)Material: High grade fully stainless steel, corrosion resistance. (3)Maximum opening of blade 45 degrees (each).(4)Sim's vaginal speculum of three different sizes. (5)Approx. Dimensions:</p> <ul style="list-style-type: none"> <li>• Medium: 70x32/75x35mm.</li> </ul> <p>(6)Handle thickness approx. 2.5mm. (7) Workmanship: All surfaces shall be free from burrs, pits, cracks. Edges shall be smoothly rounded off &amp; shall not be sharp.(8)It shall be autoclaveable/sterilizeable.</p>	Piece	1				

  
 रमकलदेव यादव  
 स्वास्थ्य संयोजक




58	Speculum sims Vaginal small	(1)Stainless steel, reusable Sim's vaginal speculum, double ended.(2)Material: High grade fully stainless steel, corrosion resistance. (3)Maximum opening of blade 45 degrees (each).(4)Sim's vaginal speculum of three different sizes. (5)Approx. Dimensions: • Small: 65x26/72x30mm. (6)Handle thickness approx. 2.5mm. (7) Workmanship: All surfaces shall be free from burrs, pits, cracks. Edges shall be smoothly rounded off & shall not be sharp.(8)It shall be autoclaveable/sterilizeable.	Piece	1					
59	Speculum Cusco's Vaginal large	(1)Stainless steel, reusable Cusco's vaginal speculum, easy to use. (2) Cusco's Vaginal speculum of three different sizes. (3) Material: High grade fully stainless steel, corrosion resistance. (4)Approx. Dimensions: • Large: 110x37mm (5)Maximum opening of blade 45 deg. (each).(6) Handle thickness approx. 2.5mm.(7) Blade size must be appropriate, confirming to sizes mentioned above. (8)Workmanship: All surfaces shall be polished and free from burrs, pits, cracks. Edges shall be smoothly rounded off & shall not be sharp.(9)It shall be autoclaveable/sterilizeable.							
60	Speculum Cusco's Vaginal medium	(1)Stainless steel, reusable Cusco's vaginal speculum, easy to use. (2) Cusco's Vaginal speculum of three different sizes. (3) Material: High grade fully stainless steel, corrosion resistance. (4)Approx. Dimensions: • Medium: 90 x 36 mm (5)Maximum opening of blade 45 deg. (each).(6) Handle thickness approx. 2.5mm.(7) Blade size must be appropriate, confirming to sizes mentioned above. (8)Workmanship: All surfaces shall be polished and free from burrs, pits, cracks. Edges shall be smoothly rounded off & shall not be sharp.(9)It shall be autoclaveable/sterilizeable.							
61	Speculum Cusco's Vaginal small	(1)Stainless steel, reusable Cusco's vaginal speculum, easy to use. (2) Cusco's Vaginal speculum of three different sizes. (3) Material: High grade fully stainless steel, corrosion resistance. (4)Approx. Dimensions: • Small: 80 x 24 mm (5)Maximum opening of blade 45 deg. (each).(6) Handle thickness approx. 2.5mm.(7) Blade size must be appropriate, confirming to sizes mentioned above. (8)Workmanship: All surfaces shall be polished and free from burrs, pits, cracks. Edges shall be smoothly rounded off & shall not be sharp.(9)It shall be autoclaveable/sterilizeable.	Piece	1					

  
सकलदेव  
स्वास्थ्य संयोजक



62	Laryngoscope Set child and adult	(1)Laryngoscopy to assist endotracheal intubation of adults, children and infants during anaesthesia or resuscitation. (2)Battery powered laryngoscope unit (handle to take C-size batteries). (3)It shall have large hollow, cylindrical, slightly ribbed handle. (4)Handle to be made of either chromium-plated or stainless steel. (5)It shall have provision to insert two batteries (type LR14, size C, 1.5 V). (6)It shall have provision for fitting various sizes and types of depressors. (7)Light to be activated when depressor is engaged.(8)It shall be autoclaveable /sterilizeable. (9)Shall come with a set of four stainless steel depressors, with halogen bulb: Macintosh type: • Curved Nr 2, length approx. 110 mm • Curved Nr 3, length approx. 135 mm • Curved Nr 4, length approx. 155 mm Miller type: • Straight Nr 1, length approx. 100 mm	Set	1				
63	Iv Stand Four hooks	IV Stand four hooks for birthing center	Set	1				
64	BP instruments Child	(1) Aneroid sphygmomanometer (2) Cuffs for child size (regular) (3) Inflation bulb (4) Carrying pouch (5) Packed in easy carrying high quality pouch made of waterproof cloth to accommodate cuff, and inflation bulb. (6)Gauge to be calibrated in 2 mm Hg units. (7) Must provide blood pressure cuffs for child size and for adult size (regular). (8) The child cuff is with double Velcro fastening, enabling it to be adjusted to fit tightly around the arm. (9)Dial graduated up to 300mm.Hg, with pressure release valve. (10)The bag is inflated by means of the flexible bulb connected via a tube (length approximately 60cm, flexible and with reliable quick connector).	Set	1				

  
संकलित यादव  
स्वास्थ्य संयोजक

